



## cell-BOCS: Bio-Optofluidics Cell Sorter

Glückstad, Jesper; Palima, Darwin; Bañas, Andrew Rafael; Pedersen, Finn

*Publication date:*  
2012

*Document Version*  
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

*Citation (APA):*  
Glückstad, J., Palima, D., Bañas, A. R., & Pedersen, F. (2012). *cell-BOCS: Bio-Optofluidics Cell Sorter*. Abstract from Focus on Microscopy, FOM 2012, Singapore, Indonesia. <http://www.focusonmicroscopy.org/2012/>

---

### General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

## cell-BOCS: Bio-Optofluidics Cell Sorter

Jesper Glückstad, Darwin Palima, Andrew Banas, Finn Pedersen  
DTU Fotonik, Dept. of Photonics Engineering  
Technical Univ. Denmark, DK-2800 Kgs. Lyngby, Denmark

[www.ppo.dk](http://www.ppo.dk)  
[www.optorobotix.com](http://www.optorobotix.com)

Within the framework of the recent DTU spin-out activity OptoRobotix we are developing an active cell sorter [1] that utilizes parallel microscopic machine vision for cell identification. Particles are identified based on visual features such as shape, size and color using image processing. The sorter shares features from our previously developed BioPhotonics Workstation [2, 3]. Hence, it benefits from the extended axial manipulation range provided by the low numerical aperture geometry. Detected particles are catapulted axially by several hundred microns – all with full parallelism over a large field of view, allowing them to be displaced from one laminar flow region to another. As the sorting motion is transverse to the viewing plane, multiple particles can be catapulted at the same time, therefore enabling a fully parallel sorting process [4, 5]. The cell-BOCS is developed with small footprint such that it can operate as a table top device, a significant advantage over flow cytometry or FACS systems.



**The first commercial prototype of the cell-BOCS promoted by DTU spin-out OptoRobotix**

- [1] J. Glückstad, *Nature Materials* 3, 9-10 (2004).
- [2] J. Glückstad, *Nature Photonics* 5, 7-8 (2011).
- [3] E. Papagiakoumou, F. Anselmi, A. Begue, V. de Sars, J. Glückstad, E. Isacoff, V. Emiliani, *Nature Methods* 7, 848-854 (2010).
- [4] J. Glückstad, D. Palima, J. Dam, I. Perch-Nielsen, *Jrn. of Optics A: Pure and Appl. Opt.* 11, 034012 (2009).
- [5] I. Perch-Nielsen, D. Palima, J. Dam, and J. Glückstad, *Jrn. of Optics A: Pure and Appl. Opt.* 11, 034013 (2009).